



Product specifications

Name	Anti-h TSH 5404 SP-5			
Specificity	Antibody recognizes human thyrotropin and its free beta subunit			
Description	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components.			
Product code	100026			
Product buffer solution	0.9 % NaCl, 0.095 % NaN $_3$ as a preservative			
Shelf life and storage	36 months from manufacturing at 2–8 °C			
Subclass	IgG ₁			
Analyte description	Thyroid-stimulating hormone (also known as TSH or thyrotropin) is a peptide hormone synthesized and secreted by thyrotrope cells in the anterior pituitary gland which regulates the endocrine function of the thyroid gland. TSH levels are tested in the blood of patients suspected of suffering from excess (hyperthyroidism), or deficiency (hypothyroidism) of thyroid hormone.			

Parameters tested on each lot

Product appearance	Liquid, may turn slightly opaque during storage
Product concentration	5.0 mg/ml (+/- 10 %)
Immunoreactivity	80–120 % compared to the reference sample in an FIA test
IEF Profile	6.6–7.2
Purity	≥ 95 %

Kinetic parameters

Association rate constant	8.5 x 10 ⁵ 1/Ms
Dissociation rate constant	3.9 x 10 ^{-s} 1/s
Affinity constant	K _A = 2.2 x 10 ¹⁰ 1/M; K _D = 4.6 x 10 ⁻¹¹ M (= 0.05 nM)
Determination method	SPR analysis (ProteOn XPR36)
Determination antigen	TSH, Scripps (Cat T0114, Lot 2414402)

Oy Medix Biochemica Ab Klovinpellontie 3 FI-02180 Espoo, Finland

medix@medixbiochemica.com www.medixbiochemica.com VAT reg.no. FI14631532

Legal disclaimer Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. Medix/MAB and Medix/Antigens are trademarks of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.





Cross-reactivities

hCG < 0.05 % (Boehringer, Cat 253065, Lot 10774821-25) LH 1.0 % (Scripps Laboratories, Cat L0814, Lot 125711) FSH 1.0 % (Boehringer, Cat 252999, Lot 1483403)

Epitope

Group 1

Pair recommendations

		DETECTION					
		5401	5404	5405	5407	5408	5409
CAPTURE	5401	-	-	+	+	-	+
	5404	-	-	+	+	-	+
	5405	-	-	-	-	+	+
	5407	-	-	-	-	+	+
	5408	-	-	+	+	-	+
	5409	-	-	+	+	-	-

Following pairs are especially recommended for the below mentioned assays: CLIA: 5405 (capture) – 5409 (detection), 5407 – 5409 and 5409 – 5407 LF: 5405 (membrane) – 5409 (particles), 5407 – 5409, 5408 – 5405 5408 – 5407, 5408 – 5409, 5409 – 5407

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested FIA, LF, CLIA

Native TSH antigen Lee Biosolutions 996-50 and 996-51.

TEMPERATURE, TIME	RESULT
-70 °C, 21 days	Not Determined (N/D)
-20 °C, 21 days	N/D
+4 °C, 21 days	N/D
+35 °C, 7 days	N/D
+35 °C, 21 days	N/D
+45 °C, 3 days	N/D
+45 °C, 7 days	N/D

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

Miscellaneous

Antigens tested

Product stability

References

Brockmann, E.-C., Pettersson, K and Vehniäinen M. (2010) Use of high-capacity surface with oriented recombinant antibody fragments in a 5-min immunoassay for thyroid-stimulating hormone. Anal. Biochem., 396:242-249

Eskola, J., Mäkinen, P., Oksa, L., Loikas, K., Nauma, M., Jiang, Q., Håkansson, M., Suomi, J. and Kulmala, S. (2006) Commpetitive immunoassay by hot electron-induced

Oy Medix Biochemica Ab Klovinpellontie 3 FI-02180 Espoo, Finland medix@medixbiochemica.com www.medixbiochemica.com VAT reg.no. FI14631532 Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. Medix/MAB and MedixAntigens are trademarks of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.





electrochemiluminescence detection and using a semiautomatic electrochemiluminometer. J. Luminesc., 118:238-244

Helenius, T. and Tikanoja, S. (1986) A sensitive and practical immunoradiometric assay of thyrotropin. Clin. Chem. 32:514-518

von Lode, P., Hagrén, V., Palenius, T. and Lövgren, T., (2003) One-step quantitative thyrotropin assay for the detection of hypothyroidsm in point-of-care conditions. Clin. Biochem., 36:121-128

Wu, F.-B., Han, S.-Q. and He, Y.-F. (2002) Time-resolved immunofluorometry of serum hTSH with enhanced sensitivity. J. Immunoass. Immunochem., 23(2):191-210

Oy Medix Biochemica Ab Klovinpellontie 3 FI-02180 Espoo, Finland medix@medixbiochemica.com www.medixbiochemica.com VAT reg.no. FI14631532 Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. Medix/MAB and MedixAntigens are trademarks of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.