

Anti-h D-Dimer 1404 SPTN-5

Product overview

| | |
|--------------------------------|--|
| Catalog number | 100479 |
| Specificity | Antibody recognizes human D-dimer |
| Description | Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components. |
| Product buffer solution | 50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN ₃ as a preservative |
| Shelf life and storage | 24 months from manufacturing at 2–8 °C |
| Subclass | IgG ₁ |
| Analyte description | D-dimer (DD) is a fibrin degradation product created during fibrinolysis when plasmin degrades the fibrin clot. In clinical diagnostics, D-dimer test can be used to exclude deep venous thrombosis (DVT), pulmonary embolism (PE) or disseminated intravascular coagulation (DIC). D-dimer is also valuable for monitoring patients during and after anticoagulant treatment for recurrent DVT. |

Parameters tested on each lot

| | |
|------------------------------|--|
| Product appearance | Liquid, may turn slightly opaque during storage |
| Product concentration | 5.0 mg/ml (+/- 10 %) |
| Immunoreactivity | 80–120 % compared to the reference sample in an FIA test |
| IEF Profile | 6.8–7.5 |
| Purity | ≥ 95 % |

Kinetic parameters

| | | |
|-----------------------------------|--|--|
| Association rate constant | 2.4 x 10 ⁵ 1/Ms | 5.4 x 10 ⁵ 1/Ms |
| Dissociation rate constant | Does not dissociate | 8.8 x 10 ⁻⁴ 1/s |
| Affinity constant | K _A = Not Applicable (N/A) K _D = Not Applicable (N/A) | K _A = 6.2 x 10 ⁸ 1/M K _D = 2.2 x 10 ⁻⁹ M (= 2.2 nM) |
| Determination method | SPR (ProteOn XPR36) | BLI (Octet RED96e) |
| Determination antigen | FDP-D-Dimer, Chrystal Chem Inc. | D-Dimer (native), Lee Biosolutions (Cat 200-09) |



Legal disclaimer

Cross-reactivities Does not recognize human fibrinogen

Epitope Not Determined (N/D)

Pair recommendations

| | | DETECTION | | | | | | | |
|---------|------|-----------|------|------|------|------|------|------|------|
| | | 1401 | 1402 | 1403 | 1404 | 1405 | 1407 | 1408 | 1409 |
| CAPTURE | 1401 | - | - | + | + | + | + | + | + |
| | 1402 | - | - | - | + | - | - | - | - |
| | 1403 | - | - | - | + | - | - | - | - |
| | 1404 | - | + | + | - | + | - | - | - |
| | 1405 | + | + | + | + | - | + | + | - |
| | 1407 | + | + | - | - | - | - | - | - |
| | 1408 | + | + | + | - | - | + | - | + |
| | 1409 | + | - | - | - | - | - | + | - |

Following pairs are especially recommended for the below mentioned assays:

FIA: 1408 (capture) – 1409 (detection), 1409 – 1408, 1401 – 1408, 1401 – 1409, and 1408 – 1401

IT: 1403 – 1404 and 1404 – 1407

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested FIA, IT

Antigens tested Native D-Dimer, Lee Biosolutions (Cat. 200-09, 200-12 and 200-13).

| TEMPERATURE, TIME | RESULT |
|-------------------|--------|
| -70 °C, 21 days | OK |
| -20 °C, 21 days | OK |
| +4 °C, 21 days | OK |
| +35 °C, 21 days | OK |
| +45 °C, 7 days | OK |

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

Miscellaneous -

References -



Legal disclaimer