

# **Rat IL-1a ELISA Kit**

Instructions for use

Catalogue numbers: 1x96 tests: 670.090.096

**For research use only**

Fast Track Your Research.....

## Table of Contents

1. Intended use .....	2
2. Introduction .....	2
2.1. Summary .....	2
2.2. Principle of the method .....	3
3. Reagents provided and reconstitution .....	4
4. Materials required but not provided .....	4
5. Storage Instructions .....	4
6. Specimen collection, processing & storage .....	4
7. Safety & precautions for use .....	5
8. Assay Preparation .....	5
8.1. Assay Design .....	6
8.2. Preparation of Wash Buffer .....	6
8.3. Preparation of Assay Buffer .....	6
8.4. Preparation of Standard .....	7
8.5. Preparation of Biotin-Conjugate .....	7
9. Method .....	8
10. Data Analysis .....	9
11. Assay limitations .....	11
12. Performance Characteristics .....	11
12.1. Sensitivity .....	12
12.2. Specificity .....	12
12.3. Precision .....	12
12.4. Dilution Parallelism .....	12
12.5. Stability .....	12
13. Assay Summary .....	12

# Rat IL-1 $\alpha$ ELISA KIT

## 1. Intended use

The rat IL-1 $\alpha$  ELISA is an enzyme-linked immunosorbent assay for the quantitative detection of rat IL-1 $\alpha$ .  
**The rat IL-1 $\alpha$  ELISA is for research use only. Not for diagnostic or therapeutic procedures.**

## 2. Introduction

### 2.1. Summary


The interleukin-1 (IL-1) species represent an important family of biologically active mononuclear cell-derived proteins which are involved in inflammatory reactions and in immune responses. Two distinct IL-1 species, IL-1 $\alpha$  and IL-1 $\beta$ , have been identified. They share similarities such as the same molecular weight, similar biological effects and the same receptors on target cells. IL-1 proteins are produced by macrophages, monocytes and various other cell types such as adult T cell leukemias, fibroblasts, epithelial or endothelial cells, neutrophils and astrocytes. Their biological properties include pyrogenicity, bone resorption, presentation of antigen to T cells and stimulation of B and T lymphocyte proliferation.

IL-1 $\alpha$  is an extracellular peptide, its activity has been demonstrated in various biological fluids.

## 2.2. Principle of the method


An anti-rat IL-1 $\alpha$  coating antibody is adsorbed onto microwells.

**Figure 1**




Rat IL-1 $\alpha$  present in the sample or standard binds to antibodies adsorbed to the microwells. A biotin-conjugated anti-rat IL-1 $\alpha$  antibody is added and binds to rat IL-1 $\alpha$  captured by the first antibody.

**Figure 2**




Following incubation unbound biotin-conjugated anti-rat IL-1 $\alpha$  antibody is removed during a wash step. Streptavidin-HRP is added and binds to the biotin-conjugated anti-rat IL-1 $\alpha$  antibody.

**Figure 3**




Following incubation unbound Streptavidin-HRP is removed during a wash step, and substrate solution reactive with HRP is added to the wells.

**Figure 4**



A coloured product is formed in proportion to the amount of rat IL-1 $\alpha$  present in the sample or standard. The reaction is terminated by addition of acid and absorbance is measured at 450 nm. A standard curve is prepared from 7 rat IL-1 $\alpha$  standard dilutions and rat IL-1 $\alpha$  sample concentration determined.

**Figure 5**



### 3. Reagents provided and reconstitution

- 1 aluminium pouch with a **Microwell Plate** coated with polyclonal antibody to rat IL-1 $\alpha$
- 1 vial (70  $\mu$ l) **Biotin-Conjugate** anti-rat IL-1 $\alpha$  polyclonal antibody
- 1 vial (150  $\mu$ l) **Streptavidin-HRP**
- 2 vials rat IL-1 $\alpha$  **Standard** lyophilized, 2 ng/ml upon reconstitution
- 1 vial (5 ml) **Assay Buffer Concentrate** 20x (PBS with 1% Tween 20 and 10% BSA)
- 1 bottle (50 ml) **Wash Buffer Concentrate** 20x (PBS with 1% Tween 20)
- 1 vial (15 ml) **Substrate Solution** (tetramethyl-benzidine)
- 1 vial (15 ml) **Stop Solution** (1M Phosphoric acid)
- 4 **Adhesive Films**

### 4. Materials required but not provided

- Microtitre plate reader fitted with appropriate filters (450nm required with optional 620nm reference filter)
- Microplate washer or wash bottle
- 10, 50, 100, 200 and 1,000 $\mu$ l adjustable single channel micropipettes with disposable tips
- 50-300 $\mu$ l multi-channel micropipette with disposable tips
- Multichannel micropipette reagent reservoirs
- Distilled water
- Vortex mixer
- Miscellaneous laboratory plastic and/or glass, if possible sterile

### 5. Storage Instructions

Store kit reagents between 2° and 8°C. Immediately after use remaining reagents should be returned to cold storage (2° to 8°C). Expiry of the kit and reagents is stated on labels.

Expiry of the kit components can only be guaranteed if the components are stored properly, and if, in case of repeated use of one component, this reagent is not contaminated by the first handling.

### 6. Specimen collection, processing & storage

Cell culture supernatant and serum were tested with this assay. Other body fluids might be suitable for use in the assay. Remove serum from the clot as soon as possible after clotting.

Samples containing a visible precipitate must be clarified prior to use in the assay. Do not use grossly hemolyzed or lipemic specimens.

Samples should be aliquoted and must be stored frozen at -20°C to avoid loss of bioactive rat IL-1 $\alpha$ . If samples are to be run within 24 hours, they may be stored at 2° to 8°C. Avoid repeated freeze-thaw cycles. Prior to assay, the frozen sample should be brought to room temperature slowly and mixed gently.

**Do not thaw samples in a 37°C water bath. Do not vortex or sharply agitate samples.**

## **7. Safety & precautions for use**

- Handling of reagents, serum or plasma specimens should be in accordance with local safety procedures , e.g.CDC/NIH Health manual : " Biosafety in Microbiological and Biomedical Laboratories" 1984
- Laboratory gloves should be worn at all times
- Avoid any skin contact with H<sub>2</sub>SO<sub>4</sub> and TMB. In case of contact, wash thoroughly with water
- Do not eat, drink, smoke or apply cosmetics where kit reagents are used
- Do not pipette by mouth
- When not in use, kit components should be stored refrigerated or frozen as indicated on vials or bottles labels
- All reagents should be warmed to room temperature before use. Lyophilized standards should be discarded after use
- Once the desired number of strips has been removed, immediately reseal the bag to protect the remaining strips from deterioration
- Cover or cap all reagents when not in use
- Do not mix or interchange reagents between different lots
- Do not use reagents beyond the expiration date of the kit
- Use a clean disposable plastic pipette tip for each reagent, standard, or specimen addition in order to avoid cross contamination, for the dispensing of H<sub>2</sub>SO<sub>4</sub> and substrate solution, avoid pipettes with metal parts
- Use a clean plastic container to prepare the washing solution
- Thoroughly mix the reagents and samples before use by agitation or swirling
- All residual washing liquid must be drained from the wells by efficient aspiration or by decantation followed by tapping the plate forcefully on absorbent paper. Never insert absorbent paper directly into the wells
- The TMB solution is light sensitive. Avoid prolonged exposure to light. Also, avoid contact of the TMB solution with metal to prevent colour development. Warning TMB is toxic avoid direct contact with hands. Dispose off properly
- If a dark blue colour develops within a few minutes after preparation, this indicates that the TMB solution has been contaminated and must be discarded. Read absorbance's within 1 hour after completion of the assay
- When pipetting reagents, maintain a consistent order of addition from well-to-well. This will ensure equal incubation times for all wells
- Follow incubation times described in the assay procedure
- Dispense the TMB solution within 15 min of the washing of the microtitre plate

## **8. Assay Preparation**

**Bring all reagents to room temperature before use**

## 8.1. Assay Design

Determine the number of microwell strips required to test the desired number of samples plus appropriate number of wells needed for running zeros and standards. Each sample, standard and zero should be tested **in duplicate**. Remove sufficient Microwell Strips for testing from the aluminium pouch immediately prior to use. Return any wells not required for this assay with desiccant to the pouch. Seal tightly and return to 2-8°C storage.

**Example plate layout** (example shown for a 7 point standard curve)

	Standards		Sample Wells									
	1	2	3	4	5	6	7	8	9	10	11	12
A	1000	1000										
B	500	500										
C	250	250										
D	125	125										
E	62.5	62.5										
F	31.2	31.2										
G	Zero	Zero										
H												

*All remaining empty wells can be used to test samples in duplicate*

## 8.2. Preparation of Wash Buffer

Pour entire contents (50 ml) of the Wash Buffer Concentrate (20x) into a clean 1000 ml graduated cylinder. Bring to final volume of 1000 ml with glass-distilled or deionized water. Mix gently to avoid foaming.

Transfer to a clean wash bottle and store at 2° to 25°C. Please note that Wash Buffer (1x) is stable for 30 days.

Wash Buffer (1x) may also be prepared as needed according to the following table:

Number of Strips	Wash Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	25	475
1 - 12	50	950

## 8.3. Preparation of Assay Buffer

Pour the entire contents (5 ml) of the **Assay Buffer Concentrate** (20x) into a clean 100 ml graduated cylinder. Bring to final volume of 100 ml with distilled water. Mix gently to avoid foaming.

Store at 2° to 8°C. Please note that the Assay Buffer (1x) is stable for 30 days.

Conjugate Diluent (1x) may also be prepared as needed according to the following table:

Number of Strips	Assay Buffer Concentrate (20x) (ml)	Distilled Water (ml)
1 - 6	2.5	47.5
1 - 12	5.0	95.0

#### 8.4. Preparation of Standard

Reconstitute **rat IL-1 $\alpha$  standard** by addition of distilled water.

Reconstitution volume is stated on the label of the standard vial. Allow the reconstituted standard to sit for 10-30 minutes. Swirl or mix gently to insure complete and homogeneous solubilization (concentration of reconstituted standard = 2000 pg/ml).

Label 6 tubes, one for each standard point. S1, S2, S3, S4, S5, S6. Then prepare 1:2 serial dilutions for the standard curve as follows: Pipette 225  $\mu$ l of Sample Diluent into each tube 7. Pipette 225  $\mu$ l of reconstituted standard (concentration = 2000 pg/ml) into the first tube, labelled S1, and mix (concentration of standard 1 = 1000 pg/ml). Pipette 225  $\mu$ l of this dilution into the second tube, labelled S2, and mix thoroughly before the next transfer. Repeat serial dilutions 5 more times thus creating the points of the standard curve ranging from 1000 to 31.2pg/ml.

Assay buffer serves as blank.

**Alternatively this can be conducted directly on the microtiter plate.**

#### 8.5. Preparation of Biotin-Conjugate

**Please note that the Biotin-Conjugate should be used within 30 minutes after dilution.**

Make a 1:100 dilution of the concentrated **Biotin-Conjugate** solution with Assay buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Biotin-Conjugate (ml)	Assay Buffer (1x) (ml)
1 - 6	0.03	2.97
1 - 12	0.06	5.94

#### 8.6. Preparation of Streptavidin-HRP



**Please note that the Streptavidin-HRP should be used within 30 minutes after dilution.**

Make a 1:200 dilution of the concentrated **Streptavidin-HRP** solution with Assay Buffer (1x) in a clean plastic tube as needed according to the following table:

Number of Strips	Streptavidin-HRP (ml)	Assay Buffer (1x) (ml)
1 - 6	0.03	5.97
1 - 12	0.06	11.94

## 9. Method

**We strongly recommend that every vial is mixed thoroughly without foaming prior to use.**

**Prepare all reagents as shown in section 8.**

**Note: Final preparation of Biotin conjugate (section 8.5) and Streptavidin-HRP (section 8.6) should occur immediately before use.**

Assay Step		Details
1.	Wash	a) Dispense 0.4 ml of <b>1x washing solution</b> into each well b) Aspirate the contents of each well c) Repeat steps a and b Do not allow wells to dry before use
2.	Preparation	<b>Prepare Standard curve</b> as shown in section 8.4
3.	Addition	Add 50µl of <b>Sample diluent</b> to all sample wells
4.	Addition	Add 50µl of <b>Sample</b> in duplicate to appropriate number of wells
5.	Addition	Add 100µl of each <b>Standard and appropriate blank (Assay buffer)</b> in duplicate to appropriate number of wells
6.	Addition	Add 50µl of diluted <b>biotin-conjugate</b> to all wells
7.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>2 hour(s)</b> if available on a microplate shaker set at 200 rpm
8.	Wash	Remove the cover and wash the plate as follows: a) Aspirate the liquid from each well b) Dispense 0.4 ml of <b>1x washing solution</b> into each well c) Aspirate the contents of each well d) Repeat step b and c another two times
9.	Addition	Add 100µl of <b>Streptavidin-HRP</b> solution into all wells
10.	Incubation	Cover with a plastic plate cover and incubate at room temperature (18 to 25°C) for <b>1 hour</b> if available on a microplate shaker set at 200 rpm
11.	Wash	Repeat wash step 8.
12.	Addition	Add 100µl of ready-to-use <b>TMB Substrate Solution</b> into all wells
13.	Incubation	Incubate in the dark for <b>10 minutes*</b> at room temperature. Avoid direct exposure to light by wrapping the plate in aluminium foil.
14.	Addition	Add 100µl of <b>Stop Reagent</b> into all wells
<b>Read the absorbance</b> value of each well (immediately after step 14.) on a spectrophotometer using 450 nm as the primary wavelength and optionally 620 nm as the reference wave length (610 nm to 650 nm is acceptable).		

*\*Incubation time of the substrate solution is usually determined by the ELISA reader performance. Many ELISA readers only record absorbance up to 2.0 O.D. Therefore the colour development within individual microwells must be observed by the analyst, and the substrate reaction stopped before positive wells are no longer within recordable range*

**Note: In case of incubation without shaking the obtained O.D. values may be lower than indicated below. Nevertheless the results are still valid.**

## 10. Data Analysis

Calculate the average absorbance values for each set of duplicate standards and samples. Duplicates should be within 20 per cent of the mean value.

Create a standard curve by plotting the mean absorbance for each standard concentration on the ordinate against the rat IL-1 $\alpha$  concentration on the abscissa. Draw a best fit curve through the points of the graph (a 5-parameter curve fit is recommended).

To determine the concentration of circulating rat IL-1 $\alpha$  for each sample, first find the mean absorbance value on the ordinate and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the abscissa and read the corresponding rat IL-1 $\alpha$  concentration.


**If instructions in this protocol have been followed samples have been diluted 1:2 (50  $\mu$ l sample + 50  $\mu$ l Assay Buffer (1x)), the concentration read from the standard curve must be multiplied by the dilution factor (x 2).**

**Calculation of samples with a concentration exceeding standard 1 may result in incorrect, low rat IL-1 $\alpha$  levels. Such samples require further external predilution according to expected rat IL-1 $\alpha$  values with Assay Buffer (1x) in order to precisely quantitate the actual rat IL-1 $\alpha$  level.**

It is suggested that each testing facility establishes a control sample of known rat IL-1 $\alpha$  concentration and runs this additional control with each assay. If the values obtained are not within the expected range of the control, the assay results may be invalid.

#### Figure 6

Representative standard curve for rat IL-1 $\alpha$  ELISA. Rat IL-1 $\alpha$  was diluted in serial 2-fold steps in Assay Buffer (1x). Do not use this standard curve to derive test results. A standard curve must be run for each group of microwell strips assayed.



**Table 1**

Typical data using the rat IL-1 $\alpha$  ELISA  
 Measuring wavelength: 450 nm  
 Reference wavelength: 620 nm

Standard	Rat IL-1 $\alpha$ Concentration (pg/ml)	O.D. at 450 nm	Mean O.D. at 450 nm	C.V. (%)
1	1000.0	2.074 2.041	2.058	1.1
2	500.0	1.335 1.313	1.324	1.2
3	250.0	0.826 0.777	0.802	4.3
4	125.0	0.458 0.452	0.455	0.9
5	62.5	0.285 0.286	0.286	0.2
6	31.3	0.190 0.178	0.184	4.6
Blank	0.0	0.096 0.085	0.091	8.6

The OD values of the standard curve may vary according to the conditions of assay performance (e.g. operator, pipetting technique, washing technique or temperature effects). Furthermore shelf life of the kit may affect enzymatic activity and thus colour intensity. Values measured are still valid.

## 11. Assay limitations

Do not extrapolate the standard curve beyond the maximum standard curve point. The dose-response is non-linear in this region and good accuracy is difficult to obtain. Concentrated samples above the maximum standard concentration must be diluted with Standard diluent or with your own sample buffer to produce an OD value within the range of the standard curve. Following analysis of such samples always multiply results by the appropriate dilution factor to produce actual final concentration.

The influence of various drugs on end results has not been investigated. Bacterial or fungal contamination and laboratory cross-contamination may also cause irregular results.

Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing fresh Washing Buffer, fill with Washing Buffer as indicated for each wash cycle and do not allow wells to sit uncovered or dry for extended periods.

Disposable pipette tips, flasks or glassware are preferred, reusable glassware must be washed and thoroughly rinsed of all detergents before use.

As with most biological assays conditions may vary from assay to assay therefore **a fresh standard curve must be prepared and run for every assay.**

## 12. Performance Characteristics

## 12.1. Sensitivity

The limit of detection of rat IL-1 $\alpha$  defined as the analyte concentration resulting in an absorbance significantly higher than that of the dilution medium (mean plus 2 standard deviations) was determined to be 4 pg/ml (mean of 6 independent assays).

## 12.2. Specificity

The interference of circulating factors of the immune system was evaluated by spiking these proteins at physiologically relevant concentrations into a rat IL-1 $\alpha$  positive serum. There was no cross reactivity detected.

## 12.3. Precision

### Intra Assay

Reproducibility within the assay was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 8 serum samples containing different concentrations of rat IL-1 $\alpha$ . 2 standard curves were run on each plate. **The calculated overall intra-assay coefficient of variation was < 5%.**

### Inter-assay

Assay to assay reproducibility within one laboratory was evaluated in 3 independent experiments. Each assay was carried out with 6 replicates of 8 serum samples containing different concentrations of rat IL-1 $\alpha$ . 2 standard curves were run on each plate. **The calculated overall inter-assay coefficient of variation was < 10%.**

## 12.4. Dilution Parallelism

4 serum samples with different levels of rat IL-1 $\alpha$  were analysed at serial 2 fold dilutions with 4 replicates each. **The overall mean recovery was 112%.**

## 12.5. Stability


### Freeze-Thaw Stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C and thawed 5 times, and the rat IL-1 $\alpha$  levels determined. There was no significant loss of rat IL-1 $\alpha$  immunoreactivity detected by freezing and thawing.

### Storage Stability

Aliquots of serum samples (spiked or unspiked) were stored at -20°C, 2-8°C, room temperature (RT) and at 37°C, and the rat IL-1 $\alpha$  level determined after 24 h. There was no significant loss of rat IL-1 $\alpha$  immunoreactivity detected during storage at -20°C, 2-8°C and RT. A significant loss of rat IL-1 $\alpha$  immunoreactivity was detected during storage at 37°C after 24 h.

## 13. Assay Summary



**Supplier:**

**Diaclone SAS**  
**6 Rue Dr Jean-François-Xavier Girod**  
**25000 Besançon**  
**France**  
**Tel +33 (0)3 81 41 38 38**  
**Email: [techsupport@medixbiochemica.com](mailto:techsupport@medixbiochemica.com)**