

2022-09-19

## Anti-h CYFRA21-1 1603 SPTNZ-5

**Product overview** 

Catalog number 100221

**Specificity** Antibody recognizes human CYFRA 21-1 fragment of cytokeratin 19

**Description** Monoclonal mouse antibody, cultured *in vitro* under conditions free from

animal-derived components.

Product buffer solution 50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.05 % Sulfobetaine, 0.095 %

NaN<sub>3</sub> as a preservative

Shelf life and storage 12 months from manufacturing at 2–8 °C

Subclass IgG<sub>2a</sub>

Analyte description CYFRA 21-1 is a fragment of cytokeratin 19 which is expressed in various

types of epithelial cells and tumor cells of epithelial origin. Serum CYFRA 21-1 has been used as a tumor marker of different origins and especially in

non-small cell lung cancer.

Parameters tested on each lot

Product appearance Liquid, may turn slightly opaque during storage

Product concentration 5.0 mg/ml (+/- 10%)

**Immunoreactivity** 80–120% compared to the reference sample in an FIA test

**IEF Profile** 6.9–8.6

**Purity**  $\geq 95\%$ 

Kinetic parameters

**Association rate constant** 1 x 10<sup>4</sup> 1/Ms

**Dissociation rate constant** 4 x 10<sup>-5</sup> 1/s

Affinity constant  $K_A = 3 \times 10^8 \text{ 1/M}, K_D = 4 \times 10^{-9} \text{ M}$ 

**Determination method** SPR analysis (ProteOn XPR36)

**Determination antigen** Recombinant human Cytokeratin 19, Biodesign (Cat A44214H, Lot

12F15608)





## Product specification ANTIBODY

2022-09-19

Cross-reactivities 100 % cytokeratin 19

**Epitope** Not Determined (N/D)

Pair recommendations

		DETECTION			
		1602	1603	1604	1605
CAPTURE	1602	-	-	+	+
	1603	-	-	+	+
	1604	+	+	-	-
	1605	+	+	-	-

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested FIA

Antigens tested Native CYFRA 21-1 antigen, Lee Biosolutions 192-20

Product stability TEMPERATURE, TIME RESULT

-70 °C, 21 days OK
-20 °C, 21 days OK
+4 °C, 21 days OK
+25 °C, 21 days OK
+35 °C, 21 days OK
+45 °C, 7 days OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

Miscellaneous -

References -

