

## Anti-h D-Dimer 1409 SPTN-5

### Product overview

---

<b>Catalog number</b>	100800
<b>Specificity</b>	Antibody recognizes human D-dimer
<b>Description</b>	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components.
<b>Product buffer solution</b>	50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN <sub>3</sub> as a preservative
<b>Shelf life and storage</b>	24 months from manufacturing at 2–8 °C
<b>Subclass</b>	IgG <sub>1</sub>
<b>Analyte description</b>	D-dimer (DD) is a fibrin degradation product created during fibrinolysis when plasmin degrades the fibrin clot. In clinical diagnostics, D-dimer test can be used to exclude deep venous thrombosis (DVT), pulmonary embolism (PE) or disseminated intravascular coagulation (DIC). D-dimer is also valuable for monitoring patients during and after anticoagulant treatment for recurrent DVT.

### Parameters tested on each lot

---

<b>Product appearance</b>	Liquid, may turn slightly opaque during storage
<b>Product concentration</b>	5.0 mg/ml (+/- 10 %)
<b>Immunoreactivity</b>	80–120 % compared to the reference sample in an FIA test
<b>IEF Profile</b>	6.1–6.8
<b>Purity</b>	≥ 95 %

### Kinetic parameters

---

<b>Association rate constant</b>	$3.5 \times 10^5$ 1/Ms
<b>Dissociation rate constant</b>	Does not dissociate
<b>Affinity constant</b>	K <sub>A</sub> = Not Applicable (N/A) K <sub>D</sub> = Not Applicable (N/A)
<b>Determination method</b>	BLI (Octet RED96e)
<b>Determination antigen</b>	D-Dimer (native), Lee Biosolutions (Cat 200-09)

**Legal disclaimer**

**Cross-reactivities** Does not recognize human fibrinogen

**Epitope** Not Determined (N/D)

**Pair recommendations**

		DETECTION							
		1401	1402	1403	1404	1405	1407	1408	1409
CAPTURE	1401	-	-	+	+	+	+	+	+
	1402	-	-	-	+	-	-	-	-
	1403	-	-	-	+	-	-	-	-
	1404	-	+	+	-	+	-	-	-
	1405	+	+	+	+	-	+	+	-
	1407	+	+	-	-	-	-	-	-
	1408	+	+	+	-	-	+	-	+
	1409	+	-	-	-	-	-	+	-

Following pairs are especially recommended for the below mentioned assays:

FIA: 1408 (capture) – 1409 (detection), 1409 – 1408, 1401 – 1408, 1401 – 1409, and 1408 – 1401

IT: 1403 – 1404 and 1404 – 1407

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

**Platforms tested** FIA, IT

**Antigens tested** Native D-Dimer, Lee Biosolutions (Cat. 200-09, 200-12 and 200-13).

TEMPERATURE, TIME	RESULT
-70 °C, 21 days	OK
-20 °C, 21 days	OK
+4 °C, 21 days	OK
+35 °C, 21 days	OK
+45 °C, 7 days	OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

**Miscellaneous** -

**References** -



**Legal disclaimer**