Medix Biochemica

Anti-h LH 5302 SP-1

| Product overview | | | | | | |
|-------------------------------|---|--|--|--|--|--|
| Catalog number | 100018 | | | | | |
| Specificity | Antibody recognizes human luteinizing hormone (lutropin), and its beta- subunit | | | | | |
| Description | Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components. | | | | | |
| Product buffer solution | 0.9 % NaCl, 0.095 % NaN $_3$ as a preservative | | | | | |
| Shelf life and storage | 36 months from manufacturing at 2–8 °C | | | | | |
| Subclass | lgG1 | | | | | |
| Analyte description | In both males and females, LH is essential for reproduction. In females FSH initiates follicular growth and at the time of the maturation of the follicle the estrogen rise leads to a release of LH over a 24–48 hour period. This 'LH surge' triggers ovulation thereby not only releasing the egg, but also initiating the conversion of the residual follicle into a corpus luteum that, in turn, produces progesterone to prepare the endometrium for a possible implantation. LH is necessary to maintain luteal function for the first two weeks. In case of a pregnancy luteal function will be further maintained by the action of hCG from the newly established pregnancy. In the male, LH acts upon the Leydig cells of the testis and is responsible for the production of testosterone. | | | | | |
| Parameters tested on each lot | | | | | | |
| Product appearance | Liquid, may turn slightly opaque during storage | | | | | |
| Product concentration | 1.0 mg/ml (+/-10 %) | | | | | |
| Immunoreactivity | 80–120 % compared to the reference sample in an FIA test | | | | | |
| IEF Profile | 5.6–6.3 | | | | | |
| Purity | ≥ 95 % | | | | | |
| Kinetic parameters | | | | | | |
| Association rate constant | 1.3 x 10 ⁶ 1/Ms | | | | | |
| Dissociation rate constant | 2.6 x 10⁻⁵ 1/s | | | | | |
| Affinity constant | $K_A = 4.8 \times 10^{10} \text{ 1/M}; K_D = 2.1 \times 10^{-11} \text{ M} (= 21 \text{ pM})$ | | | | | |
| Determination method | SPR analysis (ProteOn XPR36) | | | | | |
| Determination antigen | LH, Scripps Laboratories (Cat L0815) | | | | | |



Oy Medix Biochemica Ab Klovinpellontie 3, FI-02180 Espoo, Finland

medix@medixbiochemica.com www.medixbiochemica.com VAT reg. no. FI14631532

Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored, and used according to the instructions. Medix Biochemica's products may not be used or reproduced without Medix Biochemica's written permission.

Medix Biochemica

| Cross-reactivities | LH α | 10 % (Scripps Laboratories, Cat L0914, Lot 698811) |
|--------------------|------|--|
| | LH β | 127 % (Scripps Laboratories, Cat L1014, Lot 237711) |
| | FSH | 3 % (Scripps Laboratories, Cat F0614, Lot 805811) |
| | hCG | 4 % (Scripps Laboratories, Cat C0714, Lot 210164) |
| | TSH | 0.02 % (Scripps Laboratories, Cat T0114, Lot 181711) |

Epitope

Beta 1 as described in Nilsson et al. (2001) Two antibodies binding to the same, or closely located epitopes, belong to the same group and hence cannot be used as a pair in a sandwich assay. Epitope group numbering does not give any detailed information where the epitope is located.

1 - 43 Pair recomn

| Pair recommendations | | | DETECTION | | | | | | |
|----------------------|--|------|-----------|------|---|--|------------------------------------|--|--|
| | | | 5301 | 5302 | 5303 | 5304 | 5501 (α subunit) | | |
| | ш | 5301 | - | + | + | + | + | | |
| | CAPTURE | 5302 | + | - | - | - | + | | |
| | APT | 5303 | + | - | - | - | - | | |
| | ن ن | 5304 | + | - | - | - | - | | |
| | Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative. | | | | | | | | |
| Platforms tested | FIA, CLIA, LF | | | | | | | | |
| Antigens tested | Native LH antigen, Medix Biochemica 996-31. | | | | | | | | |
| Product stability | TEMPERATURE, TIME -70 °C, 21 days -20 °C, 21 days +4 °C, 21 days +35 °C, 7 days +35 °C, 21 days +45 °C, 3 days +45 °C, 7 days Stability testing is performed in the pro- temperatures affect the antigen binding antibody. Please note that the shelf life real time stability testing at 2–8 °C in the | | | | N/D N/D N/D N/D N/D roduct buffe | etermined (er to see w or compos 1 the first p | /hether different sition of the | | |
| Miscellaneous | Nilsson et al. (2001) analyzed epitopes of 30 different LH mAbs. Antibody 5302 was classified as belonging to epitope group beta 1, recognizing intact LH, its beta subunit as well as a common variant of LH. Antibody 5302 did not cross react with TSH or FSH but a minor cross reaction with hCG was detected. | | | | | | | | |
| References | Federici M.M. Fraser R. Lundovist C. and Lankford J.C. (1982) | | | | | | | | |

References Federici, M.M., Fraser, R., Lundqvist, C., and Lankford, J.C., (1982) Production and characterization of monoclonal antibodies human lutenizing hormones. Fed. Proc., 41

> Nilsson, C., Seppälä, M., and Pettersson, K., (2001) Immunological characterization of human luteinizing hormone with special regard to a common genetic variant. J.Endocrinol. 168:10-116



Oy Medix Biochemica Ab Klovinpellontie 3, FI-02180 Espoo, Finland

medix@medixbiochemica.com www.medixbiochemica.com VAT reg. no. FI14631532

Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored, and used according to the instructions. Medix Biochemica's products may not be used or reproduced without Medix Biochemica's written permission.