

Anti-h TSH 5404 SP-5

Product overview

Catalog number	100026
Specificity	Antibody recognizes human thyrotropin and its free beta subunit
Description	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components.
Product buffer solution	0.9 % NaCl, 0.095 % NaN ₃ as a preservative
Shelf life and storage	36 months from manufacturing at 2–8 °C
Subclass	IgG ₁
Analyte description	Thyroid-stimulating hormone (also known as TSH or thyrotropin) is a peptide hormone synthesized and secreted by thyrotrope cells in the anterior pituitary gland which regulates the endocrine function of the thyroid gland. TSH levels are tested in the blood of patients suspected of suffering from excess (hyperthyroidism), or deficiency (hypothyroidism) of thyroid hormone.

Parameters tested on each lot

Product appearance	Liquid, may turn slightly opaque during storage
Product concentration	5.0 mg/ml (+/- 10 %)
Immunoreactivity	80–120 % compared to the reference sample in an FIA test
IEF Profile	6.6–7.2
Purity	≥ 95 %

Kinetic parameters

Association rate constant	8.5×10^5 1/Ms
Dissociation rate constant	3.9×10^{-5} 1/s
Affinity constant	$K_A = 2.2 \times 10^{10}$ 1/M; $K_D = 4.6 \times 10^{-11}$ M (= 0.05 nM)
Determination method	SPR analysis (ProteOn XPR36)
Determination antigen	TSH, Scripps (Cat T0114, Lot 2414402)



Legal disclaimer

Cross-reactivities hCG < 0.05 % (Boehringer, Cat 253065, Lot 10774821-25)
 LH 1.0 % (Scripps Laboratories, Cat L0814, Lot 125711)
 FSH 1.0 % (Boehringer, Cat 252999, Lot 1483403)

Epitope Group 1

Pair recommendations

		DETECTION					
		5401	5404	5405	5407	5408	5409
CAPTURE	5401	-	-	+	+	-	+
	5404	-	-	+	+	-	+
	5405	-	-	-	-	+	+
	5407	-	-	-	-	+	+
	5408	-	-	+	+	-	+
	5409	-	-	+	+	-	-

Following pairs are especially recommended for the below mentioned assays:

CLIA: 5405 (capture) – 5409 (detection), 5407 – 5409 and 5409 – 5407

LF: 5405 (membrane) – 5409 (particles), 5407 – 5409, 5408 – 5405, 5408 – 5407, 5408 – 5409, 5409 – 5407

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Platforms tested FIA, LF, CLIA

Antigens tested Native TSH antigen Medix Biochemica 996-51.

TEMPERATURE, TIME	RESULT
-70 °C, 21 days	Not Determined (N/D)
-20 °C, 21 days	N/D
+4 °C, 21 days	N/D
+35 °C, 7 days	N/D
+35 °C, 21 days	N/D
+45 °C, 3 days	N/D
+45 °C, 7 days	N/D

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

Miscellaneous -



Legal disclaimer

References

Brockmann, E.-C., Pettersson, K and Vehniäinen M. (2010) Use of high-capacity surface with oriented recombinant antibody fragments in a 5-min immunoassay for thyroid-stimulating hormone. *Anal. Biochem.*, 396:242-249

Eskola, J., Mäkinen, P., Oksa, L., Loikas, K., Nauma, M., Jiang, Q., Håkansson, M., Suomi, J. and Kulmala, S. (2006) Competitive immunoassay by hot electron-induced electrochemiluminescence detection and using a semiautomatic electrochemiluminometer. *J. Luminesc.*, 118:238-244

Helenius, T. and Tikanoja, S. (1986) A sensitive and practical immunoradiometric assay of thyrotropin. *Clin. Chem.* 32:514-518

von Lode, P., Hagrén, V., Palenius, T. and Lövgren, T., (2003) One-step quantitative thyrotropin assay for the detection of hypothyroidism in point-of-care conditions. *Clin. Biochem.*, 36:121-128

Wu, F.-B., Han, S.-Q. and He, Y.-F. (2002) Time-resolved immunofluorometry of serum hTSH with enhanced sensitivity. *J. Immunoass. Immunochem.*, 23(2):191-210



Legal disclaimer