

PRODUCT SPECIFICATIONS

Name	Anti-h CEA 5911 SP-1
Specificity	Antibody recognizes human carcinoembryonic antigen
Description	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components
Product code	100048
Product buffer solution	0.9 % NaCl, 0.095 % NaN ₃ as a preservative
Shelf life and storage	24 months from manufacturing at 2-8 °C
Analyte description	Carcinoembryonic antigen (CEA) is a glycoprotein involved in cell adhesion. It is normally produced during fetal development, but the production of CEA stops before birth. Therefore, it is not usually present in the blood of healthy adults, but elevated levels have been found in individuals with carcinomas. CEA measurement is mainly used as a tumor marker to identify recurrences after surgical resection. CEA levels may also be raised in some non-neoplastic conditions like ulcerative colitis, pancreatitis, and cirrhosis.

PARAMETERS TESTED ON EACH LOT

Product appearance	Liquid, may turn slightly opaque during storage
Product concentration	1.0 mg/ml (+/- 10 %)
Immunoreactivity	80-120 % compared to the reference sample in an FIA test
IEF Profile	6.5 - 7.5
Purity	≥ 95 %

PARAMETERS DETERMINED DURING PRODUCT DEVELOPMENT

Subclass	IgG ₁
Association rate constant	Not Determined (N/D)
Dissociation rate constant	N/D
Affinity constant	1.5 x 10 ¹⁰ l/mol
Determination method	Radioimmunoassay (RIA)
Determination antigen	CEA, BIOSCAN (Cat 100, Lot 118-1T)

Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

Oy Medix Biochemica Ab

Headquarters and manufacturing
Klovinpellontie 3
FI-02180 Espoo, Finland

medix@medixbiochemica.com
www.medixbiochemica.com
VAT reg. no. FI4631532

Cross-reactivities Cross-reactivity with human NCA-2 (non-specific cross-reacting antigen-2).

Epitope Epitope group E as described in Bhayana et al. (1989)

Pair recommendations	CAPTURE ANTIBODY	DETECTION ANTIBODY
	-	-

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Product stability	TEMPERATURE, TIME	RESULT
	-70 °C, 21 days	N/D
	-20 °C, 21 days	OK
	+4 °C, 21 days	OK
	+30 °C, 21 days	OK
	+35 °C, 7 days	OK
	+35 °C, 21 days	Failed due to fragmentation
	+45 °C, 3 days	OK
	+45 °C, 7 days	Failed due to fragmentation

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2-8 °C in the product buffer.

Miscellaneous In Bhayana et al. (1989) authors made an epitope mapping for 11 monoclonal antibodies from Medix Biochemica and they found that the antibodies bound to 5 different epitope groups (A to F) and 5911 was found to bind to epitope group E.

References Bhayana, V. and Diamandis, E.P. (1989) A double monoclonal time-resolved immunofluorometric assay of carcinoembryonic antigen in serum. Clin. Biochem., 22:433-438