

## PRODUCT SPECIFICATIONS

<b>Name</b>	Anti-h CYFRA21-1 1602 SPTNZ-5
<b>Specificity</b>	Antibody recognizes human cytokeratin 19 CYFRA21-1 fragment
<b>Description</b>	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components
<b>Product code</b>	100220
<b>Product buffer solution</b>	50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.05 % Sulfobetaine, 0.095 % NaN <sub>3</sub> as a preservative
<b>Shelf life and storage</b>	6 months from manufacturing at 2-8 °C
<b>Analyte description</b>	CYFRA21-1 is a fragment of cytokeratin 19 which is expressed in various types of epithelial cells and tumor cells of epithelial origin. Serum CYFRA21-1 has been used as a tumor marker of different origins and especially in non-small cell lung cancer.

## PARAMETERS TESTED ON EACH LOT

<b>Product appearance</b>	Liquid, may turn slightly opaque during storage
<b>Product concentration</b>	5.0 mg/ml (+/- 10 %)
<b>Immunoreactivity</b>	80-120 % compared to the reference sample in an FIA test
<b>IEF Profile</b>	6.5 - 7.1
<b>Purity</b>	≥ 95 %

## PARAMETERS DETERMINED DURING PRODUCT DEVELOPMENT

<b>Subclass</b>	IgG <sub>1</sub>
<b>Association rate constant</b>	Not Determined (N/D)
<b>Dissociation rate constant</b>	N/D
<b>Affinity constant</b>	N/D
<b>Determination method</b>	-
<b>Determination antigen</b>	-

### Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

### Oy Medix Biochemica Ab

**Headquarters and manufacturing**  
Klovinpellontie 3  
FI-02180 Espoo, Finland

medix@medixbiochemica.com  
www.medixbiochemica.com  
VAT reg. no. FI4631532

**Cross-reactivities** 100 % Cytokeratin 19

**Epitope** N/D

<b>Pair recommendations</b>	<b>CAPTURE ANTIBODY</b>	<b>DETECTION ANTIBODY</b>
	1602	1604, 1605
	1604, 1605	1602

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

<b>Product stability</b>	<b>TEMPERATURE, TIME</b>	<b>RESULT</b>
	-70 °C, 21 days	OK
	-20 °C, 21 days	OK
	+4 °C, 21 days	OK
	+25 °C, 21 days	OK
	+35 °C, 21 days	OK
	+45 °C, 7 days	OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2-8 °C in the product buffer.

**Miscellaneous** This antibody has a high tendency to precipitate.

**References** -

**Legal disclaimer**

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

**Oy Medix Biochemica Ab**

**Headquarters and manufacturing**  
Klovinpellontie 3  
FI-02180 Espoo, Finland

medix@medixbiochemica.com  
www.medixbiochemica.com  
VAT reg. no. FI4631532