

## PRODUCT SPECIFICATIONS

<b>Name</b>	Anti-h D-Dimer 1405 SPTN-5
<b>Specificity</b>	Antibody recognizes human D-Dimer
<b>Description</b>	Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components
<b>Product code</b>	100480
<b>Product buffer solution</b>	50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN <sub>3</sub> as a preservative
<b>Shelf life and storage</b>	36 months from manufacturing at 2-8 °C
<b>Analyte description</b>	D-Dimer (DD) is a fibrin degradation product created during fibrinolysis when plasmin degrades the fibrin clot. In clinical diagnostics, D-Dimer test can be used to exclude deep venous thrombosis (DVT), pulmonary embolism (PE) or disseminated intravascular coagulation (DIC). D-Dimer is also valuable for monitoring patients during and after anticoagulant treatment for recurrent DVT.

## PARAMETERS TESTED ON EACH LOT

<b>Product appearance</b>	Liquid, may turn slightly opaque during storage
<b>Product concentration</b>	5.0 mg/ml (+/- 10 %)
<b>Immunoreactivity</b>	80-120 % compared to the reference sample in an FIA test
<b>IEF Profile</b>	6.2 – 7.2
<b>Purity</b>	≥ 95 %

## PARAMETERS DETERMINED DURING PRODUCT DEVELOPMENT

<b>Subclass</b>	IgG <sub>1</sub>
<b>Association rate constant</b>	3.5 x 10 <sup>5</sup> 1/Ms
<b>Dissociation rate constant</b>	Does not dissociate
<b>Affinity constant</b>	Not Applicable (N/A)
<b>Determination method</b>	SPR analysis (ProteOn XPR36)
<b>Determination antigen</b>	FDP-D-Dimer, Chrystal Chem Inc. (Lot SDD-1)

### Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

### Oy Medix Biochemica Ab

**Headquarters and manufacturing**  
Klovinpellontie 3  
FI-02180 Espoo, Finland

medix@medixbiochemica.com  
www.medixbiochemica.com  
VAT reg. no. FI4631532

**Cross-reactivities**

No cross-reactivity against human fibrinogen, others not tested.

**Epitope**

Not Determined (N/D)

**Pair recommendations**
**CAPTURE ANTIBODY**

1405  
1401, 1404

**DETECTION ANTIBODY**

1401, 1402, 1403, 1404, 1406, 1407  
1405

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

**Product stability**
**TEMPERATURE, TIME**

-70 °C, 21 days  
-20 °C, 21 days  
+4 °C, 21 days  
+35 °C, 7 days  
+35 °C, 21 days  
+45 °C, 3 days  
+45 °C, 7 days

**RESULT**

OK  
OK  
OK  
OK  
OK, but some charge alterations  
OK  
OK, but some charge alterations

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2-8 °C in the product buffer.

**Miscellaneous**

-

**References**

-

**Legal disclaimer**

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

**Oy Medix Biochemica Ab**

**Headquarters and manufacturing**  
Klovinpellontie 3  
FI-02180 Espoo, Finland

medix@medixbiochemica.com  
www.medixbiochemica.com  
VAT reg. no. FI4631532