

PRODUCT SPECIFICATIONS

| | |
|--------------------------------|---|
| Name | Anti-h AMH 11309 SPTN-5 |
| Specificity | Antibody recognizes human anti-Müllerian hormone (AMH) |
| Description | Monoclonal mouse antibody, cultured <i>in vitro</i> under conditions free from animal-derived components |
| Product code | 100840 |
| Product buffer solution | 50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN ₃ as a preservative |
| Shelf life and storage | Unspecified, storage at 2–8 °C |
| Analyte description | Anti-Müllerian hormone (AMH) is a glycoprotein produced by the Sertoli cells of the testis and by the granulosa cells of the ovary. It is used as biomarker to assess ovarian reserve levels. |

PARAMETERS TESTED ON EACH LOT

| | |
|------------------------------|--|
| Product appearance | Liquid, may turn slightly opaque during storage |
| Product concentration | 5.0 mg/ml (+/- 10 %) |
| Immunoreactivity | 80–120 % compared to the reference sample in an FIA test |
| IEF Profile | 5.9–7.2 |
| Purity | ≥ 95 % |

PARAMETERS DETERMINED DURING PRODUCT DEVELOPMENT

| | |
|-----------------------------------|----------------------|
| Subclass | IgG ₁ |
| Association rate constant | Not Determined (N/D) |
| Dissociation rate constant | N/D |
| Affinity constant | N/D |
| Determination method | - |
| Determination antigen | - |

Legal disclaimer

Medix Biochemica products meet their specifications if transported, stored and used according to the instructions. MedixMAB is a registered trademark of Medix Biochemica and may not be used or reproduced without Medix Biochemica's written permission.

Oy Medix Biochemica Ab

Headquarters and manufacturing
Klovinpellontie 3
FI-02180 Espoo, Finland

medix@medixbiochemica.com
www.medixbiochemica.com
VAT reg. no. FI4631532

Cross-reactivities

N/D

Epitope

Amino acids Pro274-Glu279 in the N-terminal region of AMH

Pair recommendations

| | | DETECTION | | | | | |
|---------|-------|-----------|-------|-------|-------|-------|-------|
| | | 11301 | 11302 | 11303 | 11304 | 11305 | 11309 |
| CAPTURE | 11301 | - | - | + | - | - | + |
| | 11302 | - | - | - | - | - | + |
| | 11303 | + | + | - | + | - | + |
| | 11304 | - | - | + | - | - | + |
| | 11305 | + | + | + | + | - | + |
| | 11309 | + | + | + | + | - | - |

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

Product stability

| TEMPERATURE, TIME | RESULT |
|-------------------|--------|
| -70 °C, 21 days | OK |
| -20 °C, 21 days | OK |
| +4 °C, 21 days | OK |
| +35 °C, 21 days | OK |
| +45 °C, 7 days | OK |

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer.

Miscellaneous

-

References

-